

MANUAL CUTTER ECT/PAT Model CM

To prepare corrugated board samples to perform ECT and PAT tests on the Samples Crush Tester

MANUAL CUTTER ECT/PAT Model CM

APPLICABLE STANDARDS:


PAT Tests: APPITA/AS 1301.430s, TAPPI T 821, FEFCO No. 11

ECT Tests: APPITA/AS 1301.444s, ISO 3037, ISO 13821, DIN 53149, SCAN P33, TAPPI T 811, TAPPI T 823, TAPPI T 838, TAPPI T 839, FEFCO No. 8

GENERAL INFORMATION

The sample is cut manually, obtaining rectangular or square specimens depending on the chosen cut width and up to a maximum thickness of 12 mm.

- Easy to use
- Safe use
- Robust design
- Specimen thickness up to 12 mm
- Cutting width: 25mm, 32mm, 38.1mm, 50.8mm, 80mm and 100mm



EQUIPMENT DIMENSIONS: 425 x 330 x 140 mm

NET WEIGHT: 15 Kg

PACKAGING DIMENSIONS: 500 x 440 x 290 mm (W x D x H)

GROSS WEIGHT: 22 Kg